

## **DECLARATION OF CONFORMITY**

Application of Council Directives :

**EMC Directive 2004/108/EC**

**Machinery Directive 2006/42/EC**

Standards to which Conformity is Declared :

**EN61326-1:2013**

**EN61000-3-3:2008**

**EN61000-3-2:2009**

**EN61010-1:2010 (3rd Edition)**

Manufacturer's Name :

Armen Instrument SAS

Manufacturer's Address :

Z.I de Kermelin

16, Rue Ampère

F-56 890 Saint-Avé, FRANCE

Type of Equipment :

Laboratory Equipment

Model :

**"Spot Prep II" systems**

Beginning with Serial Number :

**14-09-101P**

Month and Year of Manufacture :

**August 2014**

I, the undersigned, hereby declare that the equipment specified above  
conforms to the above Directives and Standards.

Date of Issue (Day Month Year): 22 August 2014



Lénaïck DELOURME

Quality Manager

# Certificate



Certificate no.

CU 72140552 01

**License Holder:**

Armen Instrument sas  
ZI de Kermelin  
16, rue Ampère  
56890 Saint Avé  
France

**Manufacturing Plant:**

Armen Instrument sas  
ZI de Kermelin  
16, rue Ampère  
56890 Saint Avé  
France

Test report no.: USA-CW 31480623 001

Client Reference: Lenaick Delourme

Tested to: UL 61010-1:2012  
CAN/CSA-C22.2 NO. 61010-1-12

Certified Product: High Performance Liquid Chromatography

License Fee - Units

Model Designation: Spot Prep II

7

Rated Voltage: AC 110-240V, 50/60Hz  
Rated Power: 350W max.  
Protection Class: I

Appendix: 1, 1-3

7

Licensed Test mark:



Date of Issue  
(day/mo/yr)  
04/08/2014